



华南国家计量测试中心
广东省计量科学研究院

SOUTH CHINA NATIONAL CENTER OF METROLOGY
GUANGDONG INSTITUTE OF METROLOGY



中国认可
国际互认
校准
CALIBRATION
CNAS L0730

校准证书

CALIBRATION CERTIFICATE

证书编号 GDDE202300447
Certificate No.

第 1 页, 共 6 页
Page of

委托方 优利德科技(中国)股份有限公司
Client

委托方联络信息 广东省东莞市松山湖园区工业北一路6号
Contact Information

计量器具名称 绝缘电阻测试仪 (电子式绝缘电阻表)
Description

型号规格 UT513C
Model/Type

制造厂 UNI-T
Manufacturer

出厂编号 C223175330
Serial No.

设备编号 /
Equipment No.

接收日期 2023 年 06 月 08 日
Date of Receipt Y M D

结果 见校准结果
Results Show in the results of calibration

校准日期 2023 年 06 月 09 日
Date of calibration Y M D

批准人 何洪波
Approved Signatory


核 验 何洪波
Reviewed by

校 准 叶峻江
Calibrated by

证书专用章
Stamp



东莞地址: 广东省东莞市石排镇东园大道石排段152号 邮政编码: 523343
电话: (0769)22200760 传真: (0769)22692542 E-mail: yws@scmdg.com.cn
Add: No.152, Shipai Duan, Dongyuan Road, Shipai Town, Dongguan, Guangdong, China.
Post Code: 523343 Tel: (0769)22200760 Fax: (0769)22692542



说明

证书编号 GDDE202300447
Certificate No.

DIRECTIONS

第 2 页, 共 6 页
Page of

1. 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 本中心的质量管理体系符合 ISO/IEC 17025:2017 标准的要求。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the General Administration of Quality Supervision. The quality system is in accordance with ISO/IEC 17025:2017.

2. 本中心所出具的数据均可溯源至国家计量基准和/或国际单位制 (SI)。

All data issued by this laboratory are traceable to national primary standards and/or International System of Units (SI)

3. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 本院电学实验室	温度 24 °C	相对湿度 43 %
Place	Temperature	RH

4. 本次校准的技术依据:

Reference documents for the calibration:


JJG 1005-2019 电子式绝缘电阻表检定规程 V.R. of Electronic Insulating Resistance Meters

5. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号规格/测量范围	编号	证书号/有效期/溯源单位	计量特性
Name of Equipment /Model/Type/Range	Serial No.	Certificate No./Due Date /Traceability to	Metrological Characteristic
兆欧表检定装置(直流电阻箱)/Megameter Calibrator /ZX119-8/1kΩ~211GΩ	9611	GDDG202300016 /2024-02-08 /本中心	电阻:0.2级, 电压:1级

- 注: 1. 本证书校准结果只与受校准仪器有关。 The results relate only to the items calibrated.
Note: 2. 未经本机构书面批准, 不得部分复制此证书。 This certificate shall not be reproduced except in full, without the written approval of our laboratory.
3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注。 The information Client and Contact Information are provided by client, and the Manufacturer, Model/Type, Serial No. and Equipment No. are marked on the items.
4. 本次校准日期视为发布日期。 The calibration date is the date of issue of the certificate.



校准结果

RESULTS OF CALIBRATION

证书编号 GDDE202300447
Certificate No.

原始记录号 202300447
Record No.

第 3 页, 共 6 页
Page of

一、示值误差(电阻): 见表1

Error value(resistance): Shown in table 1

表1 Table 1

测试条件 Test condition	标准值 Reference Value	指示值 Indication Value	误差 Error	最大 允许误差 MPE	结论 Conclusion
(V)	(MΩ)	(MΩ)	(MΩ)	(MΩ)	P/F
250	1.00	1.00	0.00	±0.10	Pass
	2.00	2.01	0.01	±0.15	Pass
	5.00	5.00	0.00	±0.30	Pass
	10.00	9.99	-0.01	±0.55	Pass
	20.0	20.0	0.0	±1.5	Pass
	50.0	49.9	-0.1	±3.0	Pass
	100	100	0	±10	Pass
	200	200	0	±15	Pass
	500	502	2	±30	Pass
	(GΩ)	(GΩ)	(GΩ)	(GΩ)	P/F
1.000	1.000	0.998	-0.002	±0.055	Pass
	2.00	2.00	0.00	±0.15	Pass
	5.00	4.99	-0.01	±0.30	Pass
	10.0	10.0	0.0	±2.0	Pass
	20.0	20.0	0.0	±3.0	Pass
	50.0	49.4	-0.6	±5.9	Pass
	100.0	97.4	-2.6	±10.7	Pass
	(V)	(MΩ)	(MΩ)	(MΩ)	(MΩ)
500	1.00	1.00	0.00	±0.10	Pass
	2.00	2.00	0.00	±0.15	Pass
	5.00	5.02	0.02	±0.30	Pass
	10.0	10.0	0.0	±1.0	Pass
	20.0	20.0	0.0	±1.5	Pass
	50.0	50.0	0.0	±3.0	Pass
	100.0	99.8	-0.2	±5.5	Pass
	200	201	1	±15	Pass
500	502	2	±30	Pass	



校准结果

RESULTS OF CALIBRATION

证书编号 GDDE202300447
Certificate No.

原始记录号 202300447
Record No.

第 4 页, 共 6 页
Page of

续表1 Table 1

测试条件 Test condition	标准值 Reference Value	指示值 Indication Value	误差 Error	最大 允许误差 MPE	结论 Conclusion
(V) 500	(GΩ)	(GΩ)	(GΩ)	(GΩ)	P/F
	1.00	1.00	0.00	±0.10	Pass
	2.00	2.00	0.00	±0.15	Pass
	5.00	5.00	0.00	±0.30	Pass
	10.00	9.94	-0.06	±1.09	Pass
	20.0	20.1	0.1	±3.0	Pass
	50.0	49.8	-0.2	±6.0	Pass
	100.0	98.7	-1.3	±10.9	Pass
(V) 1000	(MΩ)	(MΩ)	(MΩ)	(MΩ)	P/F
	1.00	1.00	0.00	±0.10	Pass
	2.00	2.00	0.00	±0.15	Pass
	5.00	5.00	0.00	±0.30	Pass
	10.0	10.0	0.0	±1.0	Pass
	20.0	20.0	0.0	±1.5	Pass
	50.0	50.0	0.0	±3.0	Pass
	100.0	99.9	-0.1	±5.5	Pass
	200	200	0	±15	Pass
	500	500	0	±30	Pass
	(GΩ)	(GΩ)	(GΩ)	(GΩ)	P/F
	1.00	1.01	0.01	±0.10	Pass
	2.00	2.01	0.01	±0.15	Pass
	5.00	5.00	0.00	±0.30	Pass
	10.0	10.0	0.0	±2.0	Pass
	20.0	20.1	0.1	±3.0	Pass
	50.0	50.0	0.0	±6.0	Pass
	100.0	99.2	-0.8	±10.9	Pass



校准结果

RESULTS OF CALIBRATION


证书编号 GDDE202300447
Certificate No.

原始记录号 202300447
Record No.

第 5 页, 共 6 页
Page of

续表1 Table 1

测试条件 Test condition	标准值 Reference Value	指示值 Indication Value	误差 Error	最大 允许误差 MPE	结论 Conclusion
(V)	(MΩ)	(MΩ)	(MΩ)	(MΩ)	P/F
2500	1.00	1.00	0.00	±0.10	Pass
	2.00	1.99	-0.01	±0.15	Pass
	5.00	4.99	-0.01	±0.30	Pass
	10.0	10.0	0.0	±1.0	Pass
	20.0	20.1	0.1	±1.5	Pass
	50.0	50.1	0.1	±3.0	Pass
	100	100	0	±10	Pass
	200	200	0	±15	Pass
	500	501	1	±30	Pass
	(GΩ)	(GΩ)	(GΩ)	(GΩ)	P/F
	1.00	1.01	0.01	±0.10	Pass
	2.00	2.01	0.01	±0.15	Pass
	5.00	5.03	0.03	±0.30	Pass
	10.0	10.0	0.0	±1.0	Pass
	20.0	20.1	0.1	±2.5	Pass
	50.0	49.9	-0.1	±8.5	Pass
	100.0	99.8	-0.2	±16.0	Pass
(V)	(MΩ)	(MΩ)	(MΩ)	(MΩ)	P/F
5000	1.00	0.99	-0.01	±0.10	Pass
	2.00	1.99	-0.01	±0.15	Pass
	5.00	4.97	-0.03	±0.30	Pass
	10.00	9.98	-0.02	±0.55	Pass
	20.0	20.1	0.1	±1.5	Pass
	50.0	50.2	0.2	±3.0	Pass
	100	100	0	±10	Pass
	200	201	1	±15	Pass
	500	502	2	±30	Pass
	(GΩ)	(GΩ)	(GΩ)	(GΩ)	P/F
	1.00	1.00	0.00	±0.10	Pass
	2.00	2.01	0.01	±0.15	Pass



校准结果

RESULTS OF CALIBRATION

证书编号 GDDE202300447
Certificate No.

原始记录号 202300447
Record No.

第 6 页, 共 6 页
Page of

续表1 Table 1

测试条件 Test condition	标准值 Reference Value	指示值 Indication Value	误差 Error	最大 允许误差 MPE	结论 Conclusion
(V)	(GΩ)	(GΩ)	(GΩ)	(GΩ)	P/F
5000	5.00	5.04	0.04	±0.30	Pass
	10.0	10.0	0.0	±1.0	Pass
	20.0	19.9	-0.1	±2.5	Pass
	50.0	49.9	-0.1	±9.5	Pass
	100.0	99.6	-0.4	±16.9	Pass

二、开路电压（电压）：见表2

Open circuit voltage (voltage) : Shown in table 2

表2 Table 2

标准值 Reference Value	指示值 Indication Value	标准值 允许下限 Low Limit	标准值 允许上限 High Limit	结论 Conclusion
(V)	(V)	(V)	(V)	P/F
283	250	250	300	Pass
539	500	500	600	Pass
1039	1000	1000	1200	Pass
2582	2500	2500	3000	Pass
5261	5000	5000	6000	Pass

说明:

Note:

1.本次测量结果的扩展不确定度:

The Expanded Uncertainty of Measurement:

电阻: $U_{rel}=1.6\%(\leq 1G\Omega)$, $U_{rel}=3\%(\leq 10G\Omega)$, $U_{rel}=6\%(\leq 100G\Omega)$;电压: $U_{rel}=3\%$;

包含因子 $k=2$, 本证书中给出的扩展不确定度依据JJF1059.1-2012《测量不确定度评定与表示》评定, 由合成标准不确定度乘以包含概率约为95%时对应的包含因子 k 得到。

Coverage factor $k=2$, the expanded uncertainty given in this certificate is evaluated according to JJF 1059.1-2012 "Evaluation and Expression of Uncertainty in Measurement", which is obtained by multiplying the combined standard uncertainty by the coverage factor k corresponding to the coverage probability of about 95%.

2.校准结果符合性判定依据JJF1094-2002《测量仪器特性评定》第5.3.1条款和该仪器的客户要求。

Decision rules of conformity is in JJF1094-2002《Evaluation of the Characteristics of Measuring Instruments》 and customer requirements for the instrument.

3.按照所依据技术文件的规定, 建议复校时间间隔不超过1年。

According to the demand of reference document, next calibration is proposed within 1 year.