



华南国家计量测试中心
广东省计量科学研究院

SOUTH CHINA NATIONAL CENTER OF METROLOGY
GUANGDONG INSTITUTE OF METROLOGY



中国认可
国际互认
校准
CALIBRATION
CNAS L0730

校准证书

CALIBRATION CERTIFICATE

证书编号 GDDE202400025
Certificate No.

第 1 页, 共 5 页
Page of

委托方 优利德科技(中国)股份有限公司
Client

委托方联络信息 广东省东莞市松山湖园区工业北一路6号
Contact Information

计量器具名称 防爆接地电阻测试仪 (钳形接地电阻表)
Description

型号规格 UT273Ex
Model/Type

制造厂 UNI-T
Manufacturer

出厂编号 C232495219
Serial No.

设备编号 /
Equipment No.

接收日期 2024 年 01 月 05 日
Date of Receipt Y M D

结果 见校准结果
Results

校准日期 2024 年 01 月 09 日
Date of calibration Y M D

批准人 何洪波
Approved Signatory

核 验 何洪波
Reviewed by

校 准 叶峻江
Calibrated by

证书专用章
Stamp



东莞地址: 广东省东莞市石排镇东园大道石排段152号 邮政编码: 523343
电话: (0769) 22200760 传真: (0769) 22692542 E-mail: yws@scmdg.com.cn
Add: No. 152, Shipai Duan, Dongyuan Road, Shipai Town, Dongguan, Guangdong, China.
Post Code: 523343 Tel: (0769) 22200760 Fax: (0769) 22692542



说 明

证书编号 GDDE202400025
Certificate No.

DIRECTIONS

第 2 页, 共 5 页
Page of

1. 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 本中心的质量管理体系符合 ISO/IEC 17025:2017 标准的要求。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the General Administration of Quality Supervision. The quality system is in accordance with ISO/IEC 17025:2017.

2. 本中心所出具的数据均可溯源至国家计量基准和/或国际单位制(SI)。

All data issued by this laboratory are traceable to national primary standards and/or International System of Units (SI)

3. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 A4-402安规实验室 温度 (21.6~21.9) °C 相对湿度 (58~61) %
Place Temperature RH

4. 本次校准的技术依据:

Reference documents for the calibration:

JJG 1054-2009 钳形接地电阻仪检定规程 V.R. of Clamp Earth Resistance Meters

5. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号规格/测量范围 Name of Equipment /Model/Type/Range	编号 Serial No.	证书号/有效期/溯源单位 Certificate No./Due Date /Traceability to	计量特性 Metrological Characteristic
接地电阻仪检定装置(直流电阻箱) /JD-1B/(1mΩ~1kΩ)	98400	GDDG202300432 /2024-12-03 /本中心	0.1级

注: 1. 本证书校准结果只与受校准仪器有关。The results relate only to the items calibrated.

Note: 2. 未经本机构书面批准, 不得部分复制此证书。This certificate shall not be reproduced except in full, without the written approval of our laboratory.

3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注。The information Client and Contact Information are provided by client, and the Manufacturer, Model/Type, Serial No. and Equipment No. are marked on the items.

4. 本次校准日期视为发布日期。The calibration date is the date of issue of the certificate.



校准结果 RESULTS OF CALIBRATION

证书编号 GDDE202400025
Certificate No.

原始记录号 202400025
Record No.

第 3 页, 共 5 页
Page of

一、外观及通电检查:

Appearance and power-on inspection

结论

Conclusion

合格

二、分辨力:

Resolution

结论

Conclusion

合格

三、显示能力:

Display capabilities

结论

Conclusion

合格

四、偏心位置影响(电阻)见表1:

Eccentric position effects shown in table 1

表1 Table1

量程 Range (Ω)	中心位置示值 Center Value (Ω)	偏心位置示值 Eccentricity Value (Ω)	改变量 Change Value (Ω)	允许改变量 Allow Value (Ω)	结论 Conclusion
1.0~49.9	5.1	5.1	0.0	± 0.0	Pass

五、测量重复性(电阻)见表2:

Repeatability measurements shown in table 2

表2 Table2

量程 Range (Ω)	最大值 Maximum (Ω)	最小值 Minimum (Ω)	改变量 Change Value (Ω)	允许误差 Allow Value (Ω)	结论 Conclusion
1.0~49.9	9.9	9.9	0.0	± 0.0	Pass

六、报警临界电阻(电阻)见表3:

Alarm critical resistance shown in table 3

表3 Table3

量程 Range (Ω)	标准值 Reference Value (Ω)	设定值 Indication Value (Ω)	误差 Error (Ω)	允许误差 MPE (Ω)	结论 Conclusion
1.0~49.9	1.0	1.0	0.0	± 0.1	Pass
	4.0	4.0	0.0	± 0.1	Pass
	10.1	10.0	-0.1	± 0.2	Pass
	30.0	30.0	0.0	± 0.4	Pass
	100	100	0	± 3	Pass



校准结果

RESULTS OF CALIBRATION

证书编号 GDDE202400025
Certificate No.

原始记录号 202400025
Record No.

第 4 页, 共 5 页
Page of

七、示值误差（电阻）见表4:

Value Error: shown in table 4

表4 Table4

量程 Range (Ω)	标准值 Reference Value (Ω)	指示值 Indication Value (Ω)	误差 Error (Ω)	允许误差 MPE (Ω)	结论 Conclusion P/F
0.10~0.99	0.20	0.19	-0.01	± 0.01	Pass
	0.90	0.90	0.00	± 0.02	Pass
1.0~49.9	2.0	2.1	0.1	± 0.1	Pass
	10.0	9.9	-0.1	± 0.2	Pass
	20.0	19.7	-0.3	± 0.3	Pass
	30.0	30.0	0.0	± 0.4	Pass
	45.0	44.8	-0.2	± 0.5	Pass
50.0~99.5	60.0	59.5	-0.5	± 1.4	Pass
	90.0	89.5	-0.5	± 1.8	Pass
100~199	120	119	-1	± 3	Pass
	190	188	-2	± 5	Pass
200~395	250	245	-5	± 17	Pass
	350	345	-5	± 22	Pass
400~590	450	440	-10	± 54	Pass
	550	530	-20	± 63	Pass
600~880	650	620	-30	± 144	Pass
	850	780	-70	± 176	Pass
900~1200	950	860	-90	± 245	Pass
	1000	900	-100	± 255	Pass



校准结果

RESULTS OF CALIBRATION

证书编号 GDDE202400025
Certificate No.

原始记录号 202400025
Record No.

第 5 页, 共 5 页
Page of

说明:

Note:

1. 本次测量结果的扩展不确定度:

The Expanded Uncertainty of Measurement:

电阻: $U=0.02\Omega (<1\Omega)$, $U=0.2\Omega (<10\Omega)$, $U=0.6\Omega (<100\Omega)$, $U=3\Omega (\leq 1000\Omega)$

包含因子 $k=2$, 本证书中给出的扩展不确定度依据 JJF1059.1-2012 《测量不确定度评定与表示》 评定, 由合成标准不确定度乘以包含概率约为 95% 时对应的包含因子 k 得到。

Coverage factor $k=2$, the expanded uncertainty given in this certificate is evaluated according to JJF 1059.1-2012 *Evaluation and Expression of Uncertainty in Measurement*, which is obtained by multiplying the combined standard uncertainty by the coverage factor k corresponding to the coverage probability of about 95%.

2. 校准结果符合性判定依据 JJF1094-2002 《测量仪器特性评定》 第 5.3.1 条款和该仪器说明书技术要求。

Decision rules of conformity is JJF1094-2002 *Evaluation of the Characteristic of Measuring Instruments(5.3.1)* and the technical requirements in the manual.

3. 按照所依据技术文件的规定, 建议复校时间间隔不超过 1 年。 更换重要部件、 维修或对仪器性能有怀疑时, 应及时校准。

According to the demand of reference document, next calibration is proposed within 1 year. In case of replacement of important parts, maintenance or doubt on the performance of the instrument, it shall be calibrated in time.

