



校准证书

CALIBRATION CERTIFICATE

证书编号 GDDC202400137
Certificate No.

第 1 页, 共 5 页
Page of

委托方 优利德科技(中国)股份有限公司
Client

委托方联络信息 广东省东莞市松山湖园区工业北一路6号
Contact Information

计量器具名称 智能防雷元件测试仪
Description

型号规格 UT506A
Model/Type

制造厂 UNI-T
Manufacturer

出厂编号 C221625654
Serial No.

设备编号 /
Equipment No.

接收日期 2024 年 03 月 21 日
Date of Receipt Y M D

结果 见校准结果
Results

校准日期 2024 年 03 月 22 日
Date of calibration Y M D

批准人 何洪波
Approved Signatory

核 验 何洪波
Reviewed by

校 准 齐海军
Calibrated by



证书专用章
Stamp





说 明

证书编号 GDDC202400137
Certificate No.

DIRECTIONS

第 2 页, 共 5 页
Page of

1. 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 本中心的质量管理体系符合 ISO/IEC 17025:2017 标准的要求。

This laboratory is the National Legal Metrological Verification Institution in southern China set up by the General Administration of Quality Supervision. The quality system is in accordance with ISO/IEC 17025:2017.

2. 本中心所出具的数据均可溯源至国家计量基准和/或国际单位制(SI)。

All data issued by this laboratory are traceable to national primary standards and/or International System of Units (SI)

3. 校准地点、环境条件:

Place and environmental conditions of the calibration:

地点 A4-402安规实验室 温度 (23.2~23.4) °C 相对湿度 (64~67) %
Place Temperature RH

4. 本次校准的技术依据:

Reference documents for the calibration:

JJG 1005-2019 电子式绝缘电阻表检定规程 V.R. of Electronic Insulating Resistance Meters
JJF(浙) 1088-2012 压敏电阻直流参数测试仪校准规范 C.S. for VDR DC Parameter Testers

5. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号规格/测量范围 Name of Equipment /Model/Type/Range	编号 Serial No.	证书号/有效期/溯源单位 Certificate No./Due Date /Traceability to	计量特性 Metrological Characteristic
交直流数字高压表(数字高压表) /149-10A/AC/DC (0.5~10) kV	CB001455	GDDC202300689 /2024-12-24 /本中心	AC:1级;DC:0.5级
数字多用表 /8846A/DCV/ACV:1mV~1000V ;DCA/ACA:10 μA~10A;DCR:0.1 Ω~1G Ω	2224001	GDDD202301205 /2024-10-17 /本中心	MPE:DCV:±0.003%;ACV:±0.09%;DCA:±0.055%;ACA:±0.14%;DCR:±0.011%
兆欧表检定装置(直流电阻箱) /ZX119-8/1k Ω~211G Ω	9611	GDDG202400031 /2025-01-31 /本中心	电阻:0.2级,电压:1级

- 注: 1. 本证书校准结果只与受校准仪器有关。The results relate only to the items calibrated.
- Note: 2. 未经本机构书面批准, 不得部分复制此证书。This certificate shall not be reproduced except in full, without the written approval of our laboratory.
3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注。The information Client and Contact Information are provided by client, and the Manufacturer, Model/Type, Serial No. and Equipment No. are marked on the items.
4. 本次校准日期视为发布日期。The calibration date is the date of issue of the certificate.



校准结果

RESULTS OF CALIBRATION

证书编号 GDDC202400137
Certificate No.

原始记录号 202400137
Record No.

第 3 页, 共 5 页
Page of

一、测试仪的压敏电压(直流电压 U_N): 见表1

The voltage-sensitive voltage of the tester: Shown in table 1

表1 Table 1

量程	实测值	指示值	误差	允许误差	结论
Range	Measured Value	Indication Value	Error	MPE	Conclusion
(V)	(V)	(V)	(V)	(V)	P/F
2000	202	200	-2	± 7	Pass
	503	500	-3	± 13	Pass
	1006	1000	-6	± 23	Pass
	1497	1500	3	± 33	Pass
	1998	2000	2	± 43	Pass

二、直流电流: 见表2

DCA: Shown in table 2

表2 Table 2

量程	实测值	标称值	误差	允许误差	结论
Range	Measured Value	Nominal Value	Error	MPE	Conclusion
(μ A)	(μ A)	(μ A)	(μ A)	(μ A)	P/F
200	20	20	0	± 10	Pass
	50	50	0	± 11	Pass
	100	100	0	± 12	Pass
	150	150	0	± 13	Pass
	190	190	0	± 14	Pass



校准结果
RESULTS OF CALIBRATION

证书编号 GDDC202400137
Certificate No.

原始记录号 202400137
Record No.

第 4 页, 共 5 页
Page of

三、电阻：见表3

OHM: Shown in table 3

表3 Table 3

测试电压 Test Voltage (V)	标准值 Standard Value (MΩ)	指示值 Indication Value (MΩ)	误差 Error (MΩ)	允许误差 MPE (MΩ)	结论 Conclusion P/F
500	10	10	0	±10	Pass
	20	20	0	±11	Pass
	50	50	0	±12	Pass
	100	99	-1	±13	Pass
	200	199	-1	±16	Pass
	500	499	-1	±25	Pass
	1000	996	-4	±40	Pass

四、开路电压（电压）：见表2

Open circuit voltage (voltage) : Shown in table 2

表2 Table 2

标准值 Reference Value (V)	指示值 Indication Value (V)	标准值 允许下限 Low Limit (V)	标准值 允许上限 High Limit (V)	结论 Conclusion P/F
500	500	500	600	Pass

说明:

Note:

1. 本次测量结果的扩展不确定度:

The Expanded Uncertainty of Measurement:

直流电压: $U_{rel}=0.6\%$; 直流电流: $U_{rel}=0.6\%$;

电阻: $U_{rel}=1.6\%(\leq 1G\Omega)$; 电压: $U_{rel}=3\%$;

包含因子 $k=2$, 本证书中给出的扩展不确定度依据JJF1059.1-2012《测量不确定度评定与表示》评定, 由合成标准不确定度乘以包含概率约为95%时对应的包含因子 k 得到。

Coverage factor $k=2$, the expanded uncertainty given in this certificate is evaluated according to JJF 1059.1-2012 *Evaluation and Expression of Uncertainty in Measurement*, which is obtained by multiplying the combined



校准结果

RESULTS OF CALIBRATION

证书编号 GDDC202400137
Certificate No.

原始记录号 202400137
Record No.

第 5 页, 共 5 页
Page of

standard uncertainty by the coverage factor k corresponding to the coverage probability of about 95%.

2. 校准结果符合性判定依据JJF1094-2002《测量仪器特性评定》第5.3.1条款和该仪器说明书技术要求。

Decision rules of conformity is in JJF1094-2002 *Evaluation of the Characteristic of Measuring Instruments(5.3.1)* and the technical requirements in the manual.

3. 按照所依据技术文件的规定, 建议复校时间间隔不超过1年。更换重要部件、维修或对仪器性能有怀疑时, 应及时校准。

According to the demand of reference document , next calibration is proposed within 1 year . In case of replacement of important parts , maintenance or doubt on the performance of the instrument , it shall be calibrated in time .

