



校准证书

CALIBRATION CERTIFICATE

证书编号 GDDD202400301
Certificate No.

第 1 页, 共 7 页
Page of

委托方 优利德科技(中国)股份有限公司
Client

委托方联络信息 广东省东莞市松山湖园区工业北一路6号
Contact Information

计量器具名称 多功能温度校准仪(过程校验仪)
Description

型号规格 UT715
Model/Type

制造厂 UNI-T
Manufacturer

出厂编号 C214872398
Serial No.

设备编号 /
Equipment No.

接收日期 2024 年 03 月 26 日
Date of Receipt Y M D

结果 见校准结果
Results

校准日期 2024 年 04 月 01 日
Date of calibration Y M D

批准人 何洪波
Approved Signatory

核 验 何洪波
Reviewed by

校 准 郑艺成
Calibrated by

证书专用章
Stamp





说明

证书编号 GDDD202400301
Certificate No.

DIRECTIONS

第 2 页, 共 7 页
Page of

- 本中心是国家市场监督管理总局在华南地区设立的国家法定计量检定机构, 本中心的质量管理体系符合 ISO/IEC 17025:2017 标准的要求。
This laboratory is the National Legal Metrological Verification Institution in southern China set up by the General Administration of Quality Supervision. The quality system is in accordance with ISO/IEC 17025:2017.
- 本中心所出具的数据均可溯源至国家计量基准和/或国际单位制(SI)。
All data issued by this laboratory are traceable to national primary standards and/or International System of Units (SI)
- 校准地点、环境条件:
Place and environmental conditions of the calibration:
地点 A4-406恒温实验室 温度 (21.1~21.8) °C 相对湿度 (48~59) %
Place Temperature RH
- 本次校准的技术依据:
Reference documents for the calibration:
JJF 1472-2014 过程仪表校验仪校准规范 C. S. for Process Calibrators

5. 本次校准所使用的主要计量标准器具:

Major standards of measurement used in the calibration:

设备名称/型号规格/测量范围 Name of Equipment /Model/Type/Range	编号 Serial No.	证书号/有效期/溯源单位 Certificate No./Due Date /Traceability to	计量特性 Metrological Characteristic
多功能标准源/Calibrator /5720A/DCV: (0.001~1000)V;DCA:1 μ A~ 2.2A;R:1 Ω ~100 Ω	8165209	DCsy2023-01526 /2024-06-14 /中国计量科学研究院	DCV: $U_{rel}=0.00038\%$, DCA: $U_{rel}=0.0040\%$, R: $U_{rel}=0.0010\%$ ($k=2$)
数字多用表/Multimeter /8508A/DCV/ACV:1mV~1kV;D CA/ACA:1 μ A~20A;DCR:1m Ω ~2G Ω	197464469	DBS202302389 /2024-06-11 /省计量院	DCV: $U_{rel}=0.00043\%$, DCA: $U_{rel}=0.0017\%$, R: $U_{rel}=0.0010\%$, ACV: $U_{rel}=0.009\%$, ACI: $U_{rel}=0.043\%$ ($k=2$)

- 注: 1. 本证书校准结果只与受校准仪器有关。 The results relate only to the items calibrated.
2. 未经本机构书面批准, 不得部分复制此证书。 This certificate shall not be reproduced except in full, without the written approval of our laboratory.
3. “委托方”、“委托方联络信息”由委托方提供, “制造厂”、“型号规格”、“出厂编号”以及“设备编号”为仪器上标注。 The information Client and Contact Information are provided by client, and the Manufacturer, Model/Type, Serial No. and Equipment No. are marked on the items.
4. 本次校准日期视为发布日期。 The calibration date is the date of issue of the certificate.



说明

证书编号 GDDD202400301
Certificate No.

DIRECTIONS

第 3 页, 共 7 页
Page of

设备名称/型号规格/测量范围 Name of Equipment /Model/Type/Range	编号 Serial No.	证书号/有效期/溯源单位 Certificate No./Due Date /Traceability to	计量特性 Metrological Characteristic
纳伏微欧表(数字多用表) /34420A/DCV:0.1mV~100V;D CR:1mΩ~1MΩ	MY42006324	DBS202400121 /2025-01-07 /省计量院	DCV:±0.0034%, DCR:±0.0062 %
频率计(外接石英晶振标 准)/Universal Counter /53132A/频 率:0.1Hz~12GHz;时 间:0.1μs~9999s	MY40003744	GDAD202300028 /2024-07-03 /本中心	MPE:±0.0005%

— 说明页以下空白 —



校准结果
RESULTS OF CALIBRATION

证书编号 GDDD202400301
Certificate No.

原始记录号 202400301
Record No.

第 4 页, 共 7 页
Page of

一、输出部分

Output part

1. 直流电压输出: 见表1

Dc voltage output: Shown in table 1

表1 Table1

量程 Range (mV)	标准值 Reference Value (mV)	指示值 Indication Value (mV)	误差 Error (mV)	允许误差 MPE (mV)	结论 Conclusion P/F
100	10.01	10.00	-0.01	±0.10	Pass
	50.02	50.00	-0.02	±0.11	Pass
	100.01	100.00	-0.01	±0.12	Pass
(V)	(V)	(V)	(V)	(V)	P/F
1	0.1001	0.1000	-0.0001	±0.0010	Pass
	0.5000	0.5000	0.0000	±0.0011	Pass
	1.0001	1.0000	-0.0001	±0.0012	Pass
10	1.001	1.000	-0.001	±0.010	Pass
	4.999	5.000	0.001	±0.011	Pass
	10.000	10.000	0.000	±0.012	Pass

2. 直流电流输出: 见表2

Dc current output: Shown in table 2

表2 Table2

量程 Range (mA)	标准值 Reference Value (mA)	指示值 Indication Value (mA)	误差 Error (mA)	允许误差 MPE (mA)	结论 Conclusion P/F
24	3.999	4.000	0.001	±0.003	Pass
	8.000	8.000	0.000	±0.004	Pass
	12.000	12.000	0.000	±0.004	Pass
	15.999	16.000	0.001	±0.005	Pass
	20.001	20.000	-0.001	±0.006	Pass
24.001	24.000	-0.001	±0.007	Pass	



校准结果 RESULTS OF CALIBRATION

证书编号 GDDD202400301
Certificate No.

原始记录号 202400301
Record No.

第 5 页, 共 7 页
Page of

3. 频率输出: 见表3

Frequency output: Shown in table 3

表3 Table 3

量程 Range (Hz)	标准值 Reference Value (Hz)	指示值 Indication Value (Hz)	误差 Error (Hz)	允许误差 MPE (Hz)	结论 Conclusion P/F
200	20.00	20.00	0.00	± 0.01	Pass
	100.00	100.00	0.00	± 0.02	Pass
	200.00	200.00	0.00	± 0.03	Pass
2000 (kHz)	200.0	200.0	0.0	± 0.1	Pass
	1000.0	1000.0	0.0	± 0.2	Pass
	2000.0	2000.0	0.0	± 0.3	Pass
20	2.000	2.000	0.000	± 0.001	Pass
	10.000	10.000	0.000	± 0.002	Pass
	20.000	20.000	0.000	± 0.003	Pass

二、测量部分

Measuring parts

1. 直流电压测量: 见表5

DCV measurement: Shown in table 5

表5 Table5

量程 Range (mV)	标准值 Reference Value (mV)	指示值 Indication Value (mV)	误差 Error (mV)	允许误差 MPE (mV)	结论 Conclusion P/F
200	20.00	20.00	0.00	± 0.05	Pass
	100.00	100.00	0.00	± 0.07	Pass
	200.00	200.01	0.01	± 0.09	Pass
30 (V)	3.000	3.000	0.000	± 0.003	Pass
	15.000	15.000	0.000	± 0.005	Pass
	30.000	30.002	0.002	± 0.008	Pass



校准结果 RESULTS OF CALIBRATION

证书编号 GDDD202400301
Certificate No.

原始记录号 202400301
Record No.

第 6 页, 共 7 页
Page of

2. 直流电流测量: 见表6

DCA measurement: Shown in table 6

表6 Table6

量程 Range (mA)	标准值 Reference Value (mA)	指示值 Indication Value (mA)	误差 Error (mA)	允许误差 MPE (mA)	结论 Conclusion P/F
24	4.000	4.000	0.000	±0.003	Pass
	8.000	8.000	0.000	±0.004	Pass
	12.000	11.999	-0.001	±0.004	Pass
	16.000	15.999	-0.001	±0.005	Pass
	20.000	19.999	-0.001	±0.006	Pass
	24.000	23.999	-0.001	±0.007	Pass

3. 频率测量: 见表7

Frequency measurement: Shown in table 7

表7 Table7

量程 Range (Hz)	标准值 Reference Value (Hz)	指示值 Indication Value (Hz)	误差 Error (Hz)	允许误差 MPE (Hz)	结论 Conclusion P/F	
100	10.000	9.999	-0.001	±0.002	Pass	
	50.000	49.999	-0.001	±0.006	Pass	
	100.000	100.000	0.000	±0.011	Pass	
1000	200.00	199.99	-0.01	±0.03	Pass	
	500.00	499.99	-0.01	±0.06	Pass	
	1000.00	999.97	-0.03	±0.11	Pass	
(kHz)	(kHz)	(kHz)	(kHz)	(kHz)	P/F	
	10	2.0000	1.9999	-0.0001	±0.0003	Pass
		5.0000	4.9999	-0.0001	±0.0006	Pass
	10.0000	9.9997	-0.0003	±0.0011	Pass	
100	20.000	19.999	-0.001	±0.003	Pass	
	50.000	49.999	-0.001	±0.006	Pass	
	100.000	99.997	-0.003	±0.011	Pass	



校准结果 RESULTS OF CALIBRATION

证书编号 GDDD202400301
Certificate No.

原始记录号 202400301
Record No.

第 7 页, 共 7 页
Page of

说明:

Note:

1. 本次测量结果的扩展不确定度:

The Expanded Uncertainty of Measurement:

直流电压输出: $U_{rel}=0.006\%$; 直流电流输出: $U_{rel}=0.010\%$; 频率输出: $U_{rel}=0.003\%$;

直流电压测量: $U_{rel}=0.006\%$; 直流电流测量: $U_{rel}=0.010\%$; 频率测量: $U_{rel}=0.003\%$;

包含因子 $k=2$, 本证书中给出的扩展不确定度依据JJF1059.1-2012《测量不确定度评定与表示》评定, 由合成标准不确定度乘以包含概率约为95%时对应的包含因子 k 得到。

Coverage factor $k=2$, the expanded uncertainty given in this certificate is evaluated according to JJF 1059.1-2012 *Evaluation and Expression of Uncertainty in Measurement*, which is obtained by multiplying the combined standard uncertainty by the coverage factor k corresponding to the coverage probability of about 95%.

2. 校准结果符合性判定依据JJF1094-2002《测量仪器特性评定》第5.3.1条款和该仪器说明书技术要求。

Decision rules of conformity is in JJF1094-2002 *Evaluation of the Characteristic of Measuring Instruments*(5.3.1) and the technical requirements in the manual.

3. 按照所依据技术文件的规定, 建议复校时间间隔不超过1年。更换重要部件、维修或对仪器性能有怀疑时, 应及时校准。

According to the demand of reference document, next calibration is proposed within 1 year. In case of replacement of important parts, maintenance or doubt on the performance of the instrument, it shall be calibrated in time.