

# Uni-Trend Technology (China) Co., Ltd

## TEST REPORT

### SCOPE OF WORK

EMC TESTING–UTi260B+, UTi165B+

### REPORT NUMBER

241213060GZU-001

### ISSUE DATE

19-March-2025

### [REVISED DATE]

[-----]

### PAGES

37

### DOCUMENT CONTROL NUMBER

TRF-EN IEC 61326-1:2021-a

© 2023 INTERTEK



## TEST REPORT

Applicant Name & Address : Uni-Trend Technology (China) Co., Ltd  
No 6, Gong Ye Bei 1st Road, Songshan Lake National High-Tech  
Industrial Development Zone, Dongguan City, Guangdong  
Province, China  
Manufacturing Site : Same as Applicant  
Intertek Report No: 241213060GZU-001

## Test standards

EN IEC 61326-1:2021  
EN IEC 61326-2-3:2021

## Sample Description

Product : Professional Thermal Imager  
Model No. : UTi260B+, UTi165B+  
Electrical Rating : Powered by Li-ion battery: 3.6V, 5000mAh, 18Wh  
Serial No. : Not Labeled  
Date Received : 13 December 2024  
Date Test : 23 December 2024-03 January 2025  
Conducted

Prepared and Checked By

*Richard Liu*

Richard Liu  
Engineer

Approved By:

*Dean Liu*

Dean Liu  
Sr. Project Engineer

This report is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this report. Only the Client is authorized to permit copying or distribution of this report and then only in its entirety. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and test results in this report are relevant only to the sample tested. This report by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
Room101/301/401/102/202/302/402/502/602/702/802, No. 7-2, Caipin Road, Huangpu District, Guangzhou, Guangdong,  
China

## TEST REPORT

### CONTENT

<b>TEST REPORT</b>	<b>1</b>
<b>CONTENT</b>	<b>3</b>
<b>1. TEST RESULTS SUMMARY</b>	<b>4</b>
<b>2. EMC RESULTS CONCLUSION</b>	<b>5</b>
<b>3. LABORATORY MEASUREMENTS</b>	<b>6</b>
<b>4. EQUIPMENT USED DURING TEST</b>	<b>8</b>
<b>5. EMI TEST</b>	<b>11</b>
5.1 CONTINUOUS CONDUCTED DISTURBANCE VOLTAGE TEST	11
5.1.1 Block Diagram of Test Setup	11
5.1.2 Test Setup and Procedure	11
5.1.3 Test Data and curve	12
5.2 RADIATED EMISSION BELOW 1 GHZ	14
5.2.1 Block Diagram of Test Setup	14
5.2.2 Test Setup and Procedure	14
5.2.3 Test Data and Curve	15
<b>6. HARMONICS OF CURRENT</b>	<b>17</b>
6.1 BLOCK DIAGRAM OF TEST SETUP	17
6.2 TEST SETUP AND PROCEDURE	17
<b>7. FLICKER</b>	<b>18</b>
7.1 BLOCK DIAGRAM OF TEST SETUP	18
7.2 TEST SETUP AND PROCEDURE	18
<b>8. EMS TEST</b>	<b>19</b>
8.1 EN 61000-4-2(PURSUANT TO EN IEC 61326-1) ELECTROSTATIC DISCHARGE IMMUNITY	20
8.2 EN 61000-4-3(PURSUANT TO EN IEC 61326-1) RADIATED ELECTROMAGNETIC FIELD IMMUNITY	23
8.3 EN 61000-4-4(PURSUANT TO EN IEC 61326-1) ELECTRICAL FAST TRANSIENT/BURST	26
8.4 EN 61000-4-5(PURSUANT TO EN IEC 61326-1) SURGE IMMUNITY	27
8.5 EN 61000-4-6(PURSUANT TO EN IEC 61326-1) INJECTED CURRENT (0.15 MHz TO 80 MHz)	28
8.6 EN IEC 61000-4-11(PURSUANT TO EN IEC 61326-1) VOLTAGE DIPS AND INTERRUPTIONS	29
8.7 EN 61000-4-8(PURSUANT TO EN IEC 61326-1) POWER FREQUENCY MAGNETIC FIELD IMMUNITY	30
<b>9. APPENDIX I - PHOTOS OF TEST SETUP</b>	<b>31</b>
<b>10. APPENDIX II – PHOTOS OF EUT</b>	<b>35</b>

## TEST REPORT

### 1. TEST RESULTS SUMMARY

Test Item	Standard	Result
Conducted disturbance voltage at mains ports	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 55011:2016+A1:2017+A11:2020+A2:2021	Pass
Radiated emission	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 55011:2016+A1:2017+A11:2020+A2:2021	Pass
Harmonic of current	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN IEC 61000-3-2: 2019	Pass
Flicker	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 61000-3-3: 2013+A1:2019	Pass
ESD immunity	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 61000-4-2:2009	Pass
Radiated EM field immunity	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 61000-4-3:2006 +A1:2008+A2:2010	Pass
EFT immunity	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 61000-4-4:2012	Pass
Surge immunity	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 61000-4-5:2014+A1:2017	Pass
Inject current immunity	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 61000-4-6:2014	Pass
Power frequency magnetic field immunity	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN 61000-4-8:2010	N/A
Voltage dips and interruption immunity	EN IEC 61326-1, EN IEC 61326-2-3 Reference: EN IEC 61000-4-11: 2020	Pass

Remark:

1. The symbol "N/A" in above table means Not Applicable.
2. When determining the test results, measurement uncertainty of tests has been considered.
3. The EUT belonging to Class B, Group 1 equipment, as requirement by EN 55011.

## TEST REPORT

### 2. EMC RESULTS CONCLUSION

UTi260B+ and UTi165B+ have same structure and PCB except model name.  
After evaluation, Select model UTi260B+ as full testing.

We tested the Professional Thermal Imager, Model: UTi260B+, to determine if it was in compliance with the relevant EN IEC standards as marked on the Test Results Summary. We found that the unit met the requirement of EN IEC 61326-1: 2021 and EN IEC 61326-2-3: 2021 standards when tested as received. The worst case's test data was presented in this test report.

The production units are required to conform to the initial sample as received when the units are placed on the market.

## TEST REPORT

### 3. LABORATORY MEASUREMENTS

#### Configuration Information

Support Equipment:

Equipment	Model No.	Rating	Supplier
Adapter	1600A016A2	100-240~, 50/60Hz, 0.4A	Intertek
USB cable	--	--	Client

Rated Voltage and frequency under test: 230V~; 50 Hz  
 Condition of Environment: Temperature: 22~28°C  
 Relative Humidity:35~60%  
 Atmosphere Pressure:86~106kPa

#### Notes:

- The EMI measurements had been made in the operating mode produced the largest emission in the frequency band being investigated consistent with normal applications. An attempt had been made to maximize the emission by varying the configuration of the EUT.
- The EMS measurements had been made in the frequency bands being investigated, with the EUT in the most susceptible operating mode consistent with normal applications. The configuration of the test sample had been varied to achieve maximum susceptibility.
- Test Location:  
 Intertek Testing Services Shenzhen Ltd. Guangzhou Branch  
 All tests were performed at:  
 Room101/301/401/102/202/302/402/502/602/702/802, No. 7-2, Caipin Road, Huangpu District, Guangzhou, Guangdong, China  
 Except Radiated Disturbance and Radiated Susceptibility were performed at:  
 Room102/104, No 203, KeZhu Road, Science City, GETDD Guangzhou, China

#### 4. Measurement Uncertainty

No.	Item	Measurement Uncertainty
1	Conducted Emission (9 kHz -150 kHz)	2.80 dB
2	Conducted Emission (150 kHz -30 MHz)	2.23 dB
3	Conducted Emission with VP	1.77 dB
4	Conducted Emission with AAN	4.18 dB
5	Conducted Emission with CVP and CP	3.77 dB
6	Conducted Emission with CP	2.36 dB
7	Disturbance Power (30 MHz -300 MHz)	3.17 dB
8	Radiated Emission with CDNE	1.86 dB
9	Radiated Emission (9 kHz -150 kHz) LLAS	3.48 dB
10	Radiated Emission (150 kHz -30 MHz) LLAS	3.09 dB
11	Radiated Emission (9 kHz -30 MHz) Loop	3.64 dB
12	Radiated Emission (30 MHz -1 GHz)	4.26 dB
13	Radiated Emission (1 GHz -6 GHz)	4.46 dB
14	Radiated Emission (6 GHz -18 GHz)	4.96 dB
15	Radiated Emission (18 GHz -26.5 GHz)	5.16 dB
16	Radiated Emission (26.5 GHz -40 GHz)	5.16 dB

## TEST REPORT

The measurement uncertainty describes the overall uncertainty of the given measured value during the operation of the EUT.

Measurement uncertainty is calculated in accordance with CISPR16-4-2:2011+A1:2014+A2:2018.

The measurement uncertainty is given with a confidence of 95%,  $k=2$ .

Determination of the test conclusion is based on IEC Guide 115 in consideration of measurement uncertainty.

## TEST REPORT

### 4. EQUIPMENT USED DURING TEST

#### Conducted Disturbance-Mains Terminal(2)

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM031-04	EMI receiver	ESR3	R&S	04/01/2025	05/01/2024
EM006-06	LISN	ENV216	R&S	01/09/2025	02/09/2024
SA047-111	Digital Temperature-Humidity Recorder	RS210	YIJIE	20/10/2025	21/10/2024
EM004-03	EMC shield Room	8m×4m×3m	Zhongyu	03/01/2025	04/01/2024
EM031-04-01	EMC32 software (CE)	V10.01.00	R&S	N/A	N/A

#### Radiated Disturbance (30 MHz-1 GHz)

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM030-04	3m Semi-Anechoic Chamber	9×6×6 m3	ETS-LINDGREN	9/04/2025	10/04/2024
EM031-02	EMI Test Receiver (9 kHz~7 GHz)	R&S ESR7	R&S	10/11/2025	11/11/2024
EM033-01	TRILOG Super Broadband test Antenna (30MHz-3GHz)	VULB 9163	SCHWARZBECK	8/12/2025	9/12/2024
EM031-02-01	Coaxial cable	/	R&S	9/04/2025	10/04/2024
EM036-01	Common-mode absorbing clamp	CMAD 20B	TESEQ	15/07/2025	16/07/2024
SA047-118	Digital Temperature-Humidity Recorder	RS210	YIJIE	15/07/2025	16/07/2024
EM031-04-01	EMC32 software (RE/RS)	V10.01.00	R&S	N/A	N/A

#### Electrostatic Discharge

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM077-04	ESD Simulator	NSG437	TESEQ	20/08/2025	21/08/2024
SA047-176	Digital Temperature-Humidity Recorder	RS210	YIJIE	04/01/2025	05/01/2024

#### Electrical Fast Transient/Burst(1)

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM005-12	EFT Generator	NX5 b-1-300-16	EM TEST	09/04/2025	10/04/2024
EM005-12-01	iec.control	Version 7.1.4	EM TEST	N/A	N/A



## TEST REPORT

SA047-102	Digital Temperature-Humidity Recorder	RS210	YIJIE	15/07/2025	16/07/2024
-----------	---------------------------------------	-------	-------	------------	------------

### Surge(2)

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM005-09	Surge/DIP Generator	NSG3040	TESEQ	04/06/2025	05/06/2024
EM005-09-02	WIN3000	Version 1.3.2	TESEQ	N/A	N/A
SA047-102	Digital Temperature-Humidity Recorder	RS210	YIJIE	15/07/2025	16/07/2024

### Conducted Susceptibility(1)

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM046-04	Power Amplifier	CBA230M-080	TESEQ	12/12/2025	13/12/2024
EM084-02	Signal generator	SML02	R&S	17/03/2025	18/03/2024
EM046-04-03	Attenuator	6dB	RD Communication	03/03/2025	04/03/2024
EM003-02	Coupling & Decoupling Network	CDN M2 16	TESEQ	10/11/2025	11/11/2024
EM019-01-06	Attenuator	100W 6dB	RD Communication	14/03/2025	15/03/2024
EM031-04-01	EMC32 software (RE/RS)	V10.01.00	R&S	N/A	N/A
SA047-102	Digital Temperature-Humidity Recorder	RS210	YIJIE	15/07/2025	16/07/2024

### Voltage Dips and Interruptions(1)

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM005-09	Surge/DIP Generator	NSG3040	TESEQ	04/06/2025	05/06/2024
EM005-09-01	Voltage Regulator	INA6501	TESEQ	04/06/2025	05/06/2024
EM005-09-02	WIN3000	Version 1.3.2	TESEQ	N/A	N/A
SA047-102	Digital Temperature-Humidity Recorder	RS210	YIJIE	15/07/2025	16/07/2024

### Radiated Susceptibility

Equipment No.	Equipment	Model	Manufacturer	Cal. Due date (DD-MM-YYYY)	Last calibration date (DD-MM-YYYY)
EM030-04	3m Semi-Anechoic Chamber	9×6×6 m <sup>3</sup>	ETS LINDGREN	09/04/2025	10/04/2024

## TEST REPORT

EM031-01	Signal generator	SMB100A	R&S	28/10/2025	29/10/2024
EM086-11	Power meter	NRP2	R&S	10/11/2025	11/11/2024
EM086-11-01	Power sensor	NRP-Z91	R&S	10/11/2025	11/11/2024
EM046-01	Power Amplifier	80RF1000-300	MILMEGA	03/03/2025	04/03/2024
EM046-03	Power Amplifier	AS0860-75-45	MILMEGA	02/09/2025	03/09/2024
EM061-05	Log. - Per. Broadband Antenna	VULP 9118 E	SCHWARZBEC K	09/10/2025	10/10/2023
EM061-07	Stacked Log.-Per. Broadband Antenna	STLP 9149	SCHWARZBEC K	09/10/2025	10/10/2023
EM034-01	Open Switch and Control Platform	OSP120/1505.3009K12	R&S	/	/
EM031-04-01	EMC32 software (RE/RS)	V10.01.00	R&S	/	/
SA047-118	Digital Temperature-Humidity Recorder	RS210	YIJIE	15/07/2025	16/07/2024


## TEST REPORT

### 5. EMI TEST

#### 5.1 Continuous Conducted Disturbance Voltage Test

**Test Result: Pass**

##### 5.1.1 Block Diagram of Test Setup



##### 5.1.2 Test Setup and Procedure

The EUT was set to achieve the maximum emission level. The mains terminal disturbance voltage was measured with the EUT in a shielded room. The EUT was connected to AC power source through an Artificial Mains Network which provide a 50Ω linear impedance Artificial hand is used if appropriate (for handheld apparatus). The load/control terminal disturbance voltage was measured with passive voltage probe if appropriate.

The table-top EUT was placed on a 0.8m high non-metallic table above earthed ground plane(Ground Reference Plane).And for floor standing EUT, was placed on a 0.1m high non-metallic supported on GRP. The EUT keeps a distance of at least 0.4m from a vertical metallic surface. The Artificial Mains Network is situated at a distance of 0.8m from the EUT.

During the test, mains lead of EUT excess 0.8m was folded back and forth parallel to the lead so as to form a horizontal bundle with a length between 0.3m and 0.4m.

The bandwidth of test receiver was set at 9 kHz. The frequency range from 150 kHz to 30 MHz was checked.

## TEST REPORT


### 5.1.3 Test Data and curve

At mains terminal:

Tested Wire: Live

Operation Mode: Charging and thermal imaging mode

Full Spectrum



## Final\_Result

Frequency (MHz)	QuasiPeak (dBuV)	CAverage (dBuV)	Limit (dBuV)	Margin (dB)	Meas. Time (ms)	Bandwidth (kHz)	Line	Filter	Corr. (dB)
0.530000	37.59	---	56.00	18.41	1000.0	9.000	L1	ON	9.6
0.534000	---	34.99	46.00	11.01	1000.0	9.000	L1	ON	9.6
0.886000	---	26.91	46.00	19.09	1000.0	9.000	L1	ON	9.6

Remark:


1. Corr. (dB) = LISN Factor (dB) + Cable Loss (dB)
2. Level (dBμV) = Corr. (dB) + Read Level (dBμV)
3. Delta Limit (dB) = Limit (dBμV) - Level (dBμV)

## TEST REPORT

Tested Wire: Neutral

Operation Mode: Charging and thermal imaging mode

Full Spectrum



## Final Result

Frequency (MHz)	QuasiPeak (dBuV)	CAverage (dBuV)	Limit (dBuV)	Margin (dB)	Meas. Time (ms)	Bandwidth (kHz)	Line	Filter	Corr. (dB)
0.530000	38.37	---	56.00	17.63	1000.0	9.000	N	ON	9.5
0.534000	---	26.44	46.00	19.56	1000.0	9.000	N	ON	9.5

Remark:


1. Corr. (dB) = LISN Factor (dB) + Cable Loss (dB)
2. Level (dBμV) = Corr. (dB) + Read Level (dBμV)
3. Delta Limit (dB) = Limit (dBμV)- Level (dBμV)

## TEST REPORT

### 5.2 Radiated Emission below 1 GHz

Test Result: Pass

#### 5.2.1 Block Diagram of Test Setup



#### 5.2.2 Test Setup and Procedure


The measurement was applied in a semi-anechoic chamber. The EUT and simulators were placed on a 0.8m high foamed table above the horizontal metal ground plane. The turn table rotated 360 degrees to determine the position of the maximum emission level. The EUT was set 3 meters away from the receiving antenna which was mounted on an antenna mask. The antenna moved up and down between from 1meter to 4 meters to find out the maximum emission level.

Broadband antenna was used as receiving antenna. Both horizontal and vertical polarization of the antenna was set on measurement. In order to find the maximum emission, all of the interface cables were manipulated according to EN 55011 requirement during radiated test. The bandwidth setting on Test Receiver was 120 kHz. The frequency range from 30 MHz to 1000 MHz was checked

## TEST REPORT

### 5.2.3 Test Data and Curve

Operation Mode: Charging and thermal imaging mode  
Horizontal



### QP


Frequency (MHz)	Quasi Peak (dBuV/m)	Bandwidth (kHz)	Pol	Corr. (dB)	Margin - QPK (dB)	Limit - QPK (dBuV/m)
189.280000	35.3	120.000	H	17.7	4.7	40.0
205.680000	35.6	120.000	H	19.2	4.4	40.0
222.160000	35.7	120.000	H	19.6	4.3	40.0
436.120000	39.3	120.000	H	24.1	7.7	47.0
526.640000	39.0	120.000	H	25.9	8.0	47.0
671.960000	40.2	120.000	H	28.2	6.8	47.0

Remark:

1. Corr. (dB) = Antenna Factor (dB) + Cable Loss (dB)
2. Quasi Peak (dBμV/m) = Corr. (dB) + Read Level (dBμV)
3. Margin (dB) = Limit QPK (dBμV/m) – Quasi Peak (dBμV/m)

## TEST REPORT

Vertical



## QP

Frequency (MHz)	Quasi Peak (dBuV/m)	Bandwidth (kHz)	Pol	Corr. (dB)	Margin - QPK (dB)	Limit - QPK (dBuV/m)
543.120000	42.0	120.000	V	26.2	5.0	47.0

Remark:

1. Corr. (dB) = Antenna Factor (dB) + Cable Loss (dB)
2. Quasi Peak (dBuV/m) = Corr. (dB) + Read Level (dBuV)
3. Margin (dB) = Limit QPK (dBuV/m) – Quasi Peak (dBuV/m)




## TEST REPORT

### 6. Harmonics of current

**Test Result: Pass**

#### 6.1 Block Diagram of Test Setup



#### 6.2 Test Setup and Procedure

Harmonics of the fundamental current were measured up to 40 order harmonics using a digital power meter with an analogue output and frequency analyzer which was integrated in the harmonic & flicker test system. The measurements were carried out under steady conditions.

Remarks: This product is not defined as lighting equipment, and has rated power less than 75W, therefore, no limit apply according to EN IEC 61000-3-2.

## TEST REPORT

### 7. Flicker

**Test Result: Pass**

#### 7.1 Block Diagram of Test Setup



#### 7.2 Test Setup and Procedure

##### 7.2.1 Definition

Flicker:	impression of unsteadiness of visual sensation induced by a lighting stimulus whose luminance or spectral distribution fluctuates with time.
Pst:	Short-term flicker indicator The flicker severity evaluated over a short period (in minutes); Pst=1 is the conventional threshold of irritability
Plt:	long-term flicker indicator; the flicker severity evaluated over a long period (a few hours). Using successive Pst values.
dc:	the relative steady-state voltage change
dmax:	the maximum relative voltage change
d(t):	the value during a voltage change

##### 7.2.2 Test condition

Remarks: This apparatus is unlikely to produce significant voltage fluctuations and flicker by examination of the circuit diagram and specification of it. Therefore, it is deemed to fulfill the relevant standard without testing according to clause 6.1 of EN 61000-3-3.

## TEST REPORT

### 8. EMS TEST

#### Performance Criteria:

- Criterion A: The equipment shall continue to operate as intended during and after the test. No DEGRADATION OF PERFORMANCE or LOSS OF FUNCTION is allowed below a PERFORMANCE LEVEL specified in the user documentation, when the equipment is used as intended. In the case of applying immunity tests with continuous electromagnetic phenomena, the PERFORMANCE LEVEL may be replaced by a permissible LOSS OF PERFORMANCE which shall recover, without user intervention. A permissible LOSS OF PERFORMANCE is allowed within the PERFORMANCE LEVEL only when this information is clearly provided to the end user via documentation, such as the product user manual. No change in the operating state is allowed nor is loss of data.
- Criterion B: The equipment shall continue to operate as intended after the test. No DEGRADATION OF PERFORMANCE or LOSS OF FUNCTION is allowed below a PERFORMANCE LEVEL specified in the user documentation, when the equipment is used as intended. During the test, the equipment PERFORMANCE LEVEL may be replaced by a permissible LOSS OF PERFORMANCE if such LOSS OF PERFORMANCE is detailed in the EMC test plan. A permissible LOSS OF PERFORMANCE is allowed within the PERFORMANCE LEVEL only when this information is clearly provided to the end user via documentation, such as the product user manual. An unintended change of the operating state is allowed if self-recoverable. No loss of stored data is allowed.
- Criterion C: LOSS OF FUNCTION is allowed, provided the function is self-recoverable or can be restored by the operation of the controls. Recovery procedure shall be included in the user documentation. No permanent damage to the equipment is allowed.

#### **Operation mode of EMS test:**

Test Item	Operation mode
ESD immunity	Charging and thermal imaging mode
Radiated EM field immunity	Charging and thermal imaging mode
EFT immunity	Charging and thermal imaging mode
Surge immunity	Charging and thermal imaging mode
Inject current immunity	Charging and thermal imaging mode
Voltage dips and interruption immunity	Charging and thermal imaging mode

*Note: "N/A" means Not Applicable in below text.*


## TEST REPORT

### 8.1 EN 61000-4-2(Pursuant to EN IEC 61326-1) Electrostatic Discharge Immunity

Performance criterion: B

Test Result: Pass

#### 8.1.1 Block Diagram of Test Setup



Note: HCP means Horizontal Coupling Plane,

VCP means Vertical Coupling Plane

GRP means Ground Reference Plane

#### 8.1.2 Test Setup and Procedure

The EUT was put on a 0.8m high wooden table 0.1m high for floor standing equipment standing on the ground reference plane (GRP) 3m by 2m in size, made by iron 1.0 mm thick.

A horizontal coupling plane (HCP) 1.6m by 0.8m in size was placed on the table, and the EUT with its cables were isolated from the HCP by an insulating support thick than 0.5mm. The VCP 0.5m by 0.5m in size & HCP were constructed from the same material type & thickness as that of the GRP, and connected to the GRP via a 470kΩ resistor at each end.

The distance between EUT and any of the other metallic surface excepted the GRP, HCP & VCP was greater than 1m.

The EUT was arranged and connected according to its functional requirements.

Direct static electricity discharges were applied only to those points and surface which were accessible to personnel during normal usage.

## TEST REPORT

On each preselected points 10 times of each polarity single discharge were applied. The time interval between successive single discharges was at least 1s.

The ESD generator was held perpendicular to the surface to which the discharge was applied. The discharge return cable of the generator was kept at a distance of 0.2m whilst the discharge was being applied. During the contact discharges, the tip of the discharge electrode was touched the EUT before the discharge switch was operated. During the air discharges, the round discharge tip of the discharge electrode was approached as fast as possible to touch the EUT.

Indirect discharge was conducted to objects placed near the EUT, simulated by applying the discharges of the ESD generator to a coupling plane, in the contact discharge mode.

After each discharge, the ESD generator was removed from the EUT, the generator was then retriggered for a new single discharge. For ungrounded product, a grounded carbon fibre brush with bleeder resistors (2×470 kΩ) in the grounding cable was used after each discharge to remove remnant electrostatic voltage.

For air discharge, a minimum of 10 single air discharges were applied to the selected test point for each such area.

### 8.1.3 Test Result

#### Direct Application of ESD

##### Direct Contact Discharge

Applied Voltage (kV)	No. of Discharge for each point per polarity	Result	Discharged Points
±4	10	N/A	Accessible metal parts of the EUT Conductive substrate with coating which is not declared to be insulating

##### Direct Air Discharge

Applied Voltage (kV)	No. of Discharge for each point per polarity	Result	Discharged Points
±2,±4,±8	10	Pass	All accessible points where contact discharge cannot be applied such as Displays, Indicators light, Keyboard, Button, Switch, Knob, Air gap, Slots, Hole and so on

## TEST REPORT

### Indirect Application of ESD

Horizontal Coupling Plane under the EUT

Applied Voltage (kV)	No. of Discharge for each point per polarity	Result	Discharged Point
±4	10	Pass	At the front edge of each HCP opposite the centre point of each unit of the EUT

Vertical Coupling Plane beside the EUT

Applied Voltage (kV)	No. of Discharge for each point per polarity	Result	Discharged Point
±4	10	Pass	The centre of the vertical edge of the coupling plane


## TEST REPORT

### 8.2 EN 61000-4-3(Pursuant to EN IEC 61326-1) Radiated Electromagnetic Field Immunity

Performance criterion: A

Test Result: Pass

#### 8.2.1 Block Diagram of Test Setup



## TEST REPORT

### 8.2.2 Test Setup and Procedure

The test was conducted in a fully anechoic chamber to maintain a uniform field of sufficient dimensions with respect to the EUT, and also in order to comply with various national and international laws prohibiting interference to radio communications.

The equipment was placed in the test facility on a non-conducting table 0.8m high (for floor standing EUT, is placed on a non-conducting support 0.1m height).

For all ports connected to EUT, manufacturer specified cable type and length was used, for those cables no specification, unshielded cable applied. Wire was left exposed to the electromagnetic field for a distance of 1m from the EUT. The EUT was arranged and connected according to its functional requirements

Before testing, the intensity of the established field strength had been checked by placing the field sensor at a calibration grid point, and with the field generating antenna and cables in the same positions as used for the calibration, the forward power needed to give the calibrated field strength was measured. Spot checks was made at a number of calibration grid points over the frequency range 80MHz to 6000MHz, both polarizations was checked.

After calibration, the EUT was initially placed with one face coincident with the calibration plane.

The frequency range was swept from 80 MHz to 1000 MHz at 3V/m EM field, 1.4 GHz to 6 GHz at 3V/m EM field, with the signal 80% amplitude modulated with a 1 kHz sine-wave, pausing to adjust the r.f. signal level.

The dwell time at each frequency was 3s so as that the EUT to be exercised and be able to respond.

The step size was 1% of the fundamental with linear interpolation between calibrated points. Test was performed with the generating antenna facing each of the four sides of the EUT.



## TEST REPORT

### 8.2.3 Test Result

Frequency (MHz)	Exposed Side	Field Strength (V/m)	Result
80 to 1000	Front	3V/m (r.m.s.)	Pass
80 to 1000	Left	3V/m (r.m.s.)	Pass
80 to 1000	Rear	3V/m (r.m.s.)	Pass
80 to 1000	Right	3V/m (r.m.s.)	Pass

Frequency (GHz)	Exposed Side	Field Strength (V/m)	Result
1.4 to 6.0	Front	3V/m (r.m.s.)	Pass
1.4 to 6.0	Left	3V/m (r.m.s.)	Pass
1.4 to 6.0	Rear	3V/m (r.m.s.)	Pass
1.4 to 6.0	Right	3V/m (r.m.s.)	Pass

## TEST REPORT


### 8.3 EN 61000-4-4(Pursuant to EN IEC 61326-1) Electrical Fast Transient/Burst

Tested Port: ☒ AC power    ☐ DC power    ☐ Signal/Control

Performance criterion: B

Test Result: Pass

#### 8.3.1 Block Diagram of Test Setup



#### 8.3.2 Test Setup and Procedure

The EUT was placed on a 0.1m high wooden table, standing on the ground reference plane 3m by 2m in size, made by steel 1mm thick.

The distance between the EUT and any other of the metallic surface except the GRP was greater than 0.5m.

The mains lead excess than 0.5m was folded to avoid a flat coil and situated at a distance of 0.1m above the ground reference plane to insure the distance between the coupling device and the EUT was 0.5m.

The EUT was arranged and connected to satisfy its functional requirement and supplied by the coupling-decoupling network. Repetition Frequency was 5 kHz.

#### 8.3.3 Test Result

Level	Polarity	A.C. Power supply line and protective earth terminal	D.C. Power Lines, Signal Line & Control Line
0.5 kV	+	N/A	N/A
0.5 kV	-	N/A	N/A
1 kV	+	Pass	N/A
1 kV	-	Pass	N/A

## TEST REPORT


### 8.4 EN 61000-4-5(Pursuant to EN IEC 61326-1) Surge Immunity

Tested Port: ☒ AC power    ☐ DC power    ☐ Signal/ Control

Performance criterion: B

Test Result: Pass

#### 8.4.1 Block Diagram of Test Setup



#### 8.4.2 Test Setup and Procedure

The surge was applied to the EUT power supply terminals via the capacitive coupling network.

Decoupling networks were required in order to avoid possible adverse effects on equipment not under test that might be powered by the same lines and to provide sufficient decoupling impedance to the surge wave so that the specified wave might be developed on the lines under test.

The EUT was arranged and connected according to its functional requirements.

The EUT was placed on a 0.1m high wooden support above the GRP), supplied by the coupling-decoupling network, and arranged and connected to satisfy its functional requirement. The power cord between the EUT and the coupling/decoupling network was less than 2 meters.

#### 8.4.3 Test Result

Tested Port	Level	Result
AC power	Line to line $\pm 0.5\text{kV}$	Pass
AC power	Line to earth $\pm 0.5\text{kV}$ , $\pm 1\text{kV}$	N/A
DC power	Line to line $\pm 0.5\text{kV}$	N/A
DC power	Line to earth $\pm 0.5\text{kV}$ , $\pm 1\text{kV}$	N/A
Signal/Control	Line to earth $\pm 1\text{kV}$	N/A

## TEST REPORT


### 8.5 EN 61000-4-6(Pursuant to EN IEC 61326-1) Injected Current (0.15 MHz to 80 MHz)

Tested Port: ☒ AC power ☐ DC power ☐ Signal/Control

Performance criterion: A

Test Result: Pass

#### 8.5.1 Block Diagram of Test Setup



#### 8.5.2 Test Setup and Procedure

The EUT was placed on an insulating support of 0.1m height above a ground reference Plane, arranged and connected to satisfy its functional requirement.

All relevant cables were provided with the appropriate coupling and decoupling devices at a distance between 0.1m and 0.3m from the projected geometry of the EUT on an insulating support of 0.03m height above the ground reference plane.

Test voltage was verified before each testing though power meter combined in the RF generator with AMP.

Dwell time was set to 3s and step was set as 1% to keep sufficient response time for EUT. The frequency from 0.15 MHz to 80 MHz was checked.

#### 8.5.3 Test Result

Port	Frequency (MHz)	Level	Result
A.C. Power Lines	0.15 to 80	3V (r.m.s.)	Pass
D.C. Power Lines	0.15 to 80	1V (r.m.s.)	N/A
Signal Lines	0.15 to 80	1V (r.m.s.)	N/A
Control Lines	0.15 to 80	1V (r.m.s.)	N/A

## TEST REPORT


### 8.6 EN IEC 61000-4-11(Pursuant to EN IEC 61326-1) Voltage Dips and Interruptions

Tested Port: AC power

Performance criterion: B (only for test level of 0% $U_T$  with 0.5 cycle and 1 cycle), C

Test Result: Pass

#### 8.6.1 Block Diagram of Test Setup



#### 8.6.2 Test Setup and Procedure

The EUT was placed on an insulating support of 0.8m height, standing on a ground reference plane, and arranged and connected to satisfy its functional requirement

The test was performed with the EUT connected to the test generator with the shortest power supply cable as specified by the EUT manufacturer.

The EUT was tested for each selected combination of test level and duration with a sequence of three dips/interruptions with intervals of 10 s minimum. Each representative mode of operation was tested.

Abrupt changes in supply voltage occurred at zero crossings of the voltage.

#### 8.6.3 Test Result

Test condition				
Test Level in % $U_T$	50 Hz		60 Hz	
	Duration	Result	Duration	Result
0	0.5	Pass	0.5	Pass
0	1	Pass	1	Pass
70	25	Pass	30	Pass
0	250	Pass	300	Pass

Remark:  $U_T$  is the rated voltage for the equipment.

## TEST REPORT

### **8.7 EN 61000-4-8(Pursuant to EN IEC 61326-1) Power Frequency Magnetic Field Immunity**

Tested Port: Enclosure

Performance criterion: A

Test Result: Not Applicable

Remark: Equipment containing no Hall elements or magnetic field sensors is not susceptible to magnetic field. Hence, this equipment is deemed to fulfil the magnetic field test.

## TEST REPORT

### 9. APPENDIX I - PHOTOS OF TEST SETUP

Conducted disturbance voltage at mains ports




Radiated emission (30 MHz–1000 MHz)




## TEST REPORT

ESD Immunity




Inject current immunity






## TEST REPORT

Radiated EM field immunity



Surge and DIP Immunity



## TEST REPORT

EFT Immunity



## TEST REPORT

### 10. APPENDIX II – PHOTOS OF EUT

Outside



Outside




## TEST REPORT

Inside




Inside





## TEST REPORT

PCB view



\*\*\*\*\*End of Report\*\*\*\*\*